
Midmark M11 - v1.0.5
Total Cycles: 314

Name: RITTER2
Sterilizer ID: V2313228
08/27/2021 08:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.0 F 0.0
2:00 85.2 F 0.0
4:00 135.0 F 0.0
6:00 146.8 F 0.1
8:00 141.7 F 1.8
10:00 170.9 F 5.3
12:00 190.6 F 9.8
14:00 209.7 F 6.9
16:00 217.1 F 8.6
18:00 230.1 F 13.2
20:00 232.5 F 10.4
22:00 234.3 F 8.4
24:00 246.5 F 14.8
26:00 258.2 F 21.4
28:00 268.8 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.8
1:00 272.7 F 30.4
2:00 271.9 F 29.9
3:00 272.7 F 30.4
4:00 271.9 F 29.7

Min 271.6 F 29.6
Max 273.1 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.9 F 0.0
5:00 232.3 F 0.0
10:00 224.0 F 0.0
15:00 215.3 F 0.0
20:00 224.5 F 0.0
25:00 229.2 F 0.0
30:00 230.8 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 29:05
STERILIZING: 4:00
VENTING: 2:53
DRYING: 30:00
TOTAL CYCLE: 01:07:05
CYCLE COMPLETE

08/27/2021 09:35 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____