

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001447  
13:51 09/22/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:52  
PRESSURIZING 5:37  
STERILIZING 7:03  
136.0 C 220kPa 7:03  
136.3 C 221kPa 7:33  
135.9 C 218kPa 8:03  
136.4 C 223kPa 8:33  
136.3 C 222kPa 9:03  
136.3 C 222kPa 9:34  
136.4 C 219kPa 10:04  
136.4 C 222kPa 10:33  
VENTING 10:33  
AIR DRYING 11:30  
CYCLE COMPLETE 41:30  
-----

Digital Signature #  
C7FD92B0A95373C1  
-----