
Midmark M11 - v1.0.0
Total Cycles: 1561

Name: RITTER 1
Sterilizer ID: V2178198
01/25/2022 01:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.0 F 0.0
2:00 187.6 F 4.2
4:00 191.3 F 8.9
6:00 211.9 F 8.3
8:00 222.8 F 8.4
10:00 236.1 F 13.6
12:00 238.7 F 10.9
14:00 236.5 F 9.0
16:00 250.1 F 15.7
18:00 261.4 F 22.2
20:00 272.1 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.8
1:00 272.9 F 29.5
2:00 272.2 F 28.9
3:00 272.2 F 29.2
4:00 273.0 F 29.5

Min 271.7 F 28.7
Max 273.0 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.8 F 0.0
5:00 224.0 F 0.0
10:00 228.5 F 0.0
15:00 233.6 F 0.0
20:00 234.1 F 0.0
25:00 229.2 F 0.0
30:00 224.3 F 0.0
Drying Complete

FILLING: 1:59
HEATING: 20:34
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:59:49
CYCLE COMPLETE

01/25/2022 02:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____