
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001258
13:32 15/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:18
PRESSURIZING 4:19
STERILIZING 5:21
136.0 C 220kPa 5:21
136.1 C 221kPa 5:51
136.4 C 224kPa 6:21
136.4 C 225kPa 6:51
136.3 C 222kPa 7:21
136.4 C 225kPa 7:51
136.4 C 225kPa 8:21
136.4 C 225kPa 8:51
136.4 C 225kPa 8:51
VENTING 8:51
AIR DRYING 9:48
CYCLE COMPLETE 24:48

Digital Signature #
D7BAB473E158B53D
