
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000816
10:53 10/12/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:18
PRESSURIZING 3:25
STERILIZING 4:35
136.3 C 222kPa 4:35
136.4 C 232kPa 5:05
136.3 C 231kPa 5:35
136.2 C 229kPa 6:05
136.0 C 230kPa 6:35
136.3 C 232kPa 7:05
136.2 C 231kPa 7:35
136.0 C 229kPa 8:05
136.0 C 229kPa 8:05
VENTING 8:06
AIR DRYING 8:56
CYCLE COMPLETE 23:56

Digital Signature #
A75BD3ED72AF74BF
