
Midmark M11 - v1.0.0
Total Cycles: 1657

Name: RITTER 1
Sterilizer ID: V2178198
02/23/2022 08:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 124.5 F 0.0
2:00 134.9 F 0.7
4:00 151.6 F 2.4
6:00 177.8 F 4.8
8:00 196.6 F 8.3
10:00 212.2 F 12.5
12:00 220.7 F 5.7
14:00 232.5 F 10.1
16:00 235.5 F 9.0
18:00 242.4 F 12.2
20:00 239.9 F 10.3
22:00 252.4 F 16.5
24:00 263.4 F 22.8
26:00 272.3 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.4 F 28.7
2:00 271.7 F 28.5
3:00 272.3 F 28.6
4:00 271.7 F 28.5

Min 271.4 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.5 F 0.0
5:00 209.4 F 0.0
10:00 214.2 F 0.0
15:00 214.5 F 0.0
20:00 216.4 F 0.0
25:00 212.8 F 0.0
30:00 212.8 F 0.0

Drying Complete

FILLING: 0:36
HEATING: 26:13
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:03:55
CYCLE COMPLETE

02/23/2022 09:25 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____