
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006494
7:35 18/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:10
PRESSURIZING 3:12
STERILIZING 4:22
136.4 C 221kPa 4:22
136.4 C 223kPa 4:52
135.9 C 220kPa 5:22
136.0 C 220kPa 5:52
136.4 C 218kPa 6:22
136.4 C 221kPa 6:52
136.4 C 223kPa 7:22
136.2 C 220kPa 7:52
136.2 C 220kPa 7:52
VENTING 7:52
AIR DRYING 8:39
CYCLE COMPLETE 38:39

Digital Signature #
D2B16CAB6F2F2D11
