
Midmark M11 - v1.0.5
Total Cycles: 3491

Name: RITTER2
Sterilizer ID: V2313228
04/02/2024 09:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 67.5 F 0.0
2:00 71.4 F 0.0
4:00 106.2 F 0.0
6:00 123.8 F 0.2
8:00 143.1 F 2.8
10:00 166.4 F 5.2
12:00 181.9 F 8.2
14:00 196.3 F 11.4
16:00 208.0 F 4.1
18:00 219.2 F 6.2
20:00 229.9 F 9.9
22:00 235.5 F 10.9
24:00 239.9 F 11.3
26:00 237.5 F 9.6
28:00 250.3 F 15.7
30:00 261.0 F 21.9
32:00 271.0 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.7 F 29.3
2:00 271.5 F 29.0
3:00 272.4 F 29.3
4:00 271.4 F 28.9

Min 271.4 F 28.6
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.5 F 0.0
5:00 221.8 F 0.0
10:00 191.9 F 0.0
15:00 188.8 F 0.0
20:00 184.7 F 0.0
25:00 180.6 F 0.0
30:00 177.1 F 0.0

Drying Complete

FILLING: 0:51
HEATING: 32:43
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:11:04
CYCLE COMPLETE

04/02/2024 10:36 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____