

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1981

Name: RITTER2  
Sterilizer ID: V2313228  
01/30/2023 03:55 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 213.4 F 0.0  
2:00 202.3 F 4.7  
4:00 207.9 F 10.7  
6:00 224.3 F 6.7  
8:00 238.2 F 12.9  
10:00 239.8 F 11.3  
12:00 240.2 F 10.7  
14:00 254.6 F 18.1  
16:00 266.7 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.9  
1:00 272.1 F 29.2  
2:00 271.9 F 29.5  
3:00 271.9 F 29.1  
4:00 271.9 F 29.6  
  
Min 271.3 F 28.8  
Max 272.5 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.1 F 0.0  
5:00 230.4 F 0.0  
10:00 228.9 F 0.0  
15:00 230.0 F 0.0  
20:00 221.7 F 0.0  
25:00 229.7 F 0.0  
30:00 226.9 F 0.0  
  
Drying Complete

FILLING: 2:25  
HEATING: 17:27  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:57:06  
CYCLE COMPLETE

01/30/2023 04:52 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_