
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000321
12:48 04/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:25
PRESSURIZING 3:28
STERILIZING 4:54
136.4 C 233kPa 4:54
136.3 C 232kPa 5:24
136.4 C 233kPa 5:54
136.3 C 233kPa 6:24
136.4 C 233kPa 6:54
136.2 C 232kPa 7:24
136.4 C 232kPa 7:54
136.4 C 232kPa 8:24
136.4 C 232kPa 8:24
VENTING 8:24
AIR DRYING 9:20
CYCLE COMPLETE 24:20

Digital Signature #
4EFF23354B089025
