
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 003331
16:34 29/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CYCLE ABORTED 2:04
STOP BUTTON PRESSED
NOT STERILE
END TIME 2:04

Digital Signature #
1DD7E756AC3A9A5B
