

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5183

Name: RITTER 1  
Sterilizer ID: V2178198  
12/12/2024 01:06 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 124.5 F 0.0  
2:00 148.6 F 0.0  
4:00 147.5 F 0.0  
6:00 164.9 F 0.1  
8:00 195.2 F 0.3  
10:00 208.8 F 2.3  
12:00 223.9 F 5.9  
14:00 233.7 F 9.5  
16:00 242.2 F 12.7  
18:00 240.8 F 10.6  
20:00 235.5 F 8.3  
22:00 242.7 F 11.2  
24:00 246.7 F 13.3  
26:00 257.7 F 19.2  
28:00 267.6 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.0  
1:00 271.8 F 28.5  
2:00 272.3 F 28.7  
3:00 272.0 F 28.4  
4:00 271.5 F 28.0

Min 271.2 F 27.9  
Max 272.5 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.3 F 0.0  
5:00 214.8 F 0.0  
10:00 229.4 F 0.0  
15:00 229.5 F 0.0  
20:00 230.2 F 0.0  
25:00 228.7 F 0.0  
30:00 231.2 F 0.0  
Drying Complete

FILLING: 0:43  
HEATING: 29:35  
STERILIZING: 4:00  
VENTING: 3:47  
DRYING: 30:00  
TOTAL CYCLE: 01:08:05  
CYCLE COMPLETE

12/12/2024 02:14 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_