
Midmark M11 - v1.0.0
Total Cycles: 1304

Name: RITTER 1
Sterilizer ID: V2178198
10/27/2021 06:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.4 F 0.0
2:00 88.9 F 0.0
4:00 119.9 F 0.0
6:00 153.4 F 0.0
8:00 162.9 F 0.2
10:00 172.1 F 1.6
12:00 191.3 F 4.1
14:00 224.6 F 6.7
16:00 229.1 F 10.2
18:00 224.2 F 10.2
20:00 235.3 F 9.4
22:00 237.1 F 11.6
24:00 241.7 F 11.1
26:00 236.7 F 8.8
28:00 250.2 F 15.0
30:00 260.8 F 20.9
32:00 270.5 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 272.3 F 28.5
1:00 272.4 F 28.7
2:00 272.0 F 28.7
3:00 272.2 F 28.6
4:00 271.7 F 28.7

Min 271.6 F 28.4
Max 272.4 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.9 F 0.0
5:00 225.4 F 0.0
10:00 222.0 F 0.0
15:00 218.3 F 0.0
20:00 210.6 F 0.0
25:00 206.2 F 0.0
30:00 190.1 F 0.0

Drying Complete

FILLING: 0:55
HEATING: 33:19
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:11:32
CYCLE COMPLETE

10/27/2021 07:55 AM
ACCEPTED: X
REJECTED:
Operator: U01