
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002304
9:17 20/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:49
PRESSURIZING 1:58
STERILIZING 2:59
136.4 C 222kPa 2:59
136.3 C 224kPa 3:29
136.4 C 221kPa 3:59
136.3 C 222kPa 4:29
136.4 C 225kPa 4:59
136.3 C 224kPa 5:29
136.4 C 225kPa 5:59
136.2 C 223kPa 6:29
136.2 C 223kPa 6:29
VENTING 6:29
AIR DRYING 7:21
CYCLE COMPLETE 25:42

Digital Signature #
6CC98FAF5910A57C
