
Midmark M11 - v1.0.5
Total Cycles: 387

Name: RITTER2
Sterilizer ID: V2313228
09/23/2021 07:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.7 F 0.0
2:00 162.5 F 1.7
4:00 167.9 F 5.5
6:00 194.0 F 10.8
8:00 216.5 F 5.3
10:00 235.1 F 10.7
12:00 231.1 F 8.0
14:00 241.1 F 11.5
16:00 245.2 F 13.6
18:00 257.7 F 20.4
20:00 269.1 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.7 F 29.9
2:00 271.3 F 29.0
3:00 272.4 F 30.0
4:00 271.9 F 29.2

Min 271.2 F 29.0
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.2 F 0.0
5:00 220.3 F 0.0
10:00 171.4 F 0.0
15:00 167.7 F 0.0
20:00 161.1 F 0.0
25:00 159.9 F 0.0
30:00 156.5 F 0.0
Drying Complete

FILLING: 1:45
HEATING: 21:02
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:59:52
CYCLE COMPLETE

09/23/2021 08:49 AM
ACCEPTED: X
REJECTED:
Operator: U02