
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001296
13:42 13/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:38
PRESSURIZING 3:40
STERILIZING 4:55
136.0 C 222kPa 4:55
136.2 C 223kPa 5:25
136.3 C 225kPa 5:55
136.4 C 225kPa 6:25
136.4 C 224kPa 6:55
136.4 C 225kPa 7:25
136.4 C 223kPa 7:55
136.3 C 225kPa 8:25
136.3 C 225kPa 8:25
VENTING 8:25
AIR DRYING 9:21
CYCLE COMPLETE 15:19

Digital Signature #
B9FC536444978C46
