
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001175
13:03 20/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:49
PRESSURIZING 3:51
STERILIZING 5:31
136.4 C 225kPa 5:31
136.4 C 225kPa 6:01
136.4 C 224kPa 6:31
136.4 C 223kPa 7:01
136.1 C 223kPa 7:31
136.5 C 225kPa 8:01
136.4 C 222kPa 8:31
136.0 C 221kPa 9:01
136.0 C 221kPa 9:01
VENTING 9:01
AIR DRYING 9:56
CYCLE COMPLETE 16:47

Digital Signature #
CDA59B47E80FEE80
