
Midmark M11 - v1.0.5
Total Cycles: 3188

Name: RITTER2
Sterilizer ID: V2313228
12/06/2023 09:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.4 F 0.0
2:00 71.6 F 0.0
4:00 74.5 F 0.2
6:00 94.0 F 0.8
8:00 138.0 F 2.6
10:00 153.6 F 5.5
12:00 175.9 F 8.6
14:00 193.9 F 11.7
16:00 204.8 F 4.3
18:00 213.8 F 7.8
20:00 225.8 F 11.6
22:00 227.8 F 7.5
24:00 239.0 F 12.6
26:00 238.7 F 10.7
28:00 250.3 F 16.3
30:00 260.6 F 22.0
32:00 270.0 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.4
1:00 272.2 F 30.1
2:00 272.9 F 30.0
3:00 271.8 F 29.6
4:00 271.0 F 29.3

Min 271.0 F 29.2
Max 273.1 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.5 F 0.0
5:00 200.6 F 0.0
10:00 186.2 F 0.0
15:00 181.3 F 0.0
20:00 177.2 F 0.0
25:00 173.9 F 0.0
30:00 172.6 F 0.0

Drying Complete

FILLING: 0:49
HEATING: 33:05
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:11:30
CYCLE COMPLETE

12/06/2023 10:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____