
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001745
13:25 17/04/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:00
PRESSURIZING 2:03
STERILIZING 3:05
136.4 C 225kPa 3:05
136.4 C 226kPa 3:35
136.3 C 225kPa 4:05
136.0 C 223kPa 4:35
136.3 C 225kPa 5:05
136.4 C 226kPa 5:35
136.4 C 227kPa 6:05
136.3 C 224kPa 6:35
136.3 C 224kPa 6:35
VENTING 6:35
AIR DRYING 7:26
CYCLE COMPLETE 7:44

Digital Signature #
5FCB675894D554A6
