
Midmark M11 - v1.0.5
Total Cycles: 1769

Name: RITTER2
Sterilizer ID: V2313228
11/23/2022 01:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.9 F 0.0
2:00 204.7 F 4.1
4:00 208.6 F 9.9
6:00 225.4 F 6.8
8:00 236.1 F 12.4
10:00 239.3 F 11.1
12:00 239.7 F 10.6
14:00 254.0 F 18.0
16:00 266.4 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 273.0 F 29.9
2:00 271.9 F 29.1
3:00 272.5 F 30.0
4:00 272.7 F 29.7

Min 271.6 F 29.0
Max 273.2 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.3 F 0.0
5:00 221.8 F 0.0
10:00 234.7 F 0.0
15:00 231.8 F 0.0
20:00 222.9 F 0.0
25:00 219.2 F 0.0
30:00 218.0 F 0.0
Drying Complete

FILLING: 2:08
HEATING: 17:33
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:56:51
CYCLE COMPLETE

11/23/2022 01:58 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____