

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1960

Name: RITTER2  
Sterilizer ID: V2313228  
01/23/2023 02:58 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 184.4 F 0.0  
2:00 177.4 F 0.2  
4:00 189.1 F 2.2  
6:00 205.9 F 6.4  
8:00 221.6 F 11.1  
10:00 228.9 F 7.5  
12:00 241.5 F 13.3  
14:00 241.5 F 12.1  
16:00 241.3 F 11.5  
18:00 254.0 F 18.1  
20:00 265.1 F 24.8  
22:00 271.0 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.6  
1:00 272.7 F 29.9  
2:00 272.3 F 30.2  
3:00 272.2 F 29.6  
4:00 272.6 F 30.2

Min 271.5 F 29.5  
Max 273.0 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.7 F 0.0  
5:00 237.3 F 0.0  
10:00 231.5 F 0.0  
15:00 230.6 F 0.0  
20:00 227.2 F 0.0  
25:00 230.5 F 0.0  
30:00 227.3 F 0.0  
Drying Complete

FILLING: 1:15  
HEATING: 22:39  
STERILIZING: 4:00  
VENTING: 3:34  
DRYING: 30:00  
TOTAL CYCLE: 01:01:29  
CYCLE COMPLETE

01/23/2023 03:59 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_