
Midmark M11 - v1.0.0
Total Cycles: 4608

Name: RITTER 1
Sterilizer ID: V2178198
06/05/2024 08:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.2 F 0.0
2:00 70.5 F 0.1
4:00 81.9 F 0.3
6:00 120.4 F 1.1
8:00 157.8 F 3.4
10:00 192.0 F 6.5
12:00 195.6 F 10.4
14:00 215.2 F 7.2
16:00 218.5 F 5.4
18:00 229.5 F 8.8
20:00 234.8 F 12.8
22:00 234.0 F 8.7
24:00 244.2 F 12.3
26:00 242.9 F 11.7
28:00 254.4 F 17.6
30:00 264.5 F 23.7
32:00 271.3 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.4 F 28.9
2:00 271.3 F 28.3
3:00 272.4 F 29.0
4:00 271.5 F 28.2

Min 271.2 F 28.2
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.5 F 0.0
5:00 197.3 F 0.0
10:00 189.5 F 0.0
15:00 183.3 F 0.0
20:00 191.0 F 0.0
25:00 225.2 F 0.0
30:00 230.7 F 0.0

Drying Complete

FILLING: 0:47
HEATING: 32:04
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:10:19
CYCLE COMPLETE

06/05/2024 09:24 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____