

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3693

Name: RITTER2  
Sterilizer ID: V2313228  
06/13/2024 08:23 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.4 F 0.0  
2:00 73.4 F 0.0  
4:00 104.4 F 0.4  
6:00 106.7 F 2.1  
8:00 143.9 F 5.0  
10:00 163.9 F 8.6  
12:00 184.1 F 12.3  
14:00 203.9 F 4.4  
16:00 216.0 F 7.9  
18:00 227.8 F 11.9  
20:00 228.7 F 7.5  
22:00 240.3 F 12.6  
24:00 238.8 F 10.5  
26:00 250.8 F 16.5  
28:00 261.5 F 22.6  
30:00 271.3 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.3  
1:00 272.8 F 29.8  
2:00 271.8 F 29.5  
3:00 272.7 F 29.7  
4:00 271.7 F 29.5

Min 271.6 F 29.2  
Max 272.8 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.0 F 0.0  
5:00 197.4 F 0.0  
10:00 212.9 F 0.0  
15:00 213.6 F 0.0  
20:00 218.4 F 0.0  
25:00 222.5 F 0.0  
30:00 222.1 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 30:39  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:08:46  
CYCLE COMPLETE

06/13/2024 09:31 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_