

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3749

Name: RITTER 1  
Sterilizer ID: V2178198  
09/11/2023 11:34 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 182.8 F 0.0  
2:00 187.0 F 2.4  
4:00 199.1 F 6.7  
6:00 211.4 F 11.2  
8:00 219.7 F 5.1  
10:00 231.5 F 9.8  
12:00 236.2 F 9.7  
14:00 242.0 F 11.9  
16:00 240.2 F 10.3  
18:00 252.3 F 16.3  
20:00 262.8 F 22.5  
22:00 272.6 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.2  
1:00 272.4 F 28.9  
2:00 271.7 F 28.2  
3:00 271.8 F 28.7  
4:00 272.2 F 28.7

Min 271.3 F 28.1  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.8 F 0.0  
5:00 211.6 F 0.0  
10:00 215.7 F 0.0  
15:00 214.0 F 0.0  
20:00 209.7 F 0.0  
25:00 206.6 F 0.0  
30:00 203.9 F 0.0  
Drying Complete

FILLING: 1:37  
HEATING: 22:24  
STERILIZING: 4:00  
VENTING: 3:43  
DRYING: 30:00  
TOTAL CYCLE: 01:01:45  
CYCLE COMPLETE

09/11/2023 12:35 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_