
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000428
14:04 27/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:31
PRESSURIZING 2:38
STERILIZING 3:40
136.5 C 232kPa 3:40
136.3 C 232kPa 4:10
136.2 C 230kPa 4:40
136.5 C 233kPa 5:10
136.3 C 230kPa 5:40
136.3 C 233kPa 6:10
136.1 C 230kPa 6:40
136.2 C 233kPa 7:10
136.2 C 233kPa 7:10
VENTING 7:10
AIR DRYING 8:02
CYCLE COMPLETE 8:10

Digital Signature #
13EB0483B561CF4A
