
Midmark M11 - v1.0.5
Total Cycles: 2129

Name: RITTER2
Sterilizer ID: V2313228
03/08/2023 03:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.0 F 0.0
2:00 157.6 F 2.8
4:00 176.1 F 6.8
6:00 199.2 F 11.9
8:00 215.4 F 6.4
10:00 229.9 F 11.8
12:00 233.4 F 9.0
14:00 237.8 F 9.8
16:00 247.2 F 14.3
18:00 260.1 F 21.3
20:00 271.4 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.2
1:00 273.0 F 29.9
2:00 271.7 F 29.2
3:00 272.6 F 30.1
4:00 272.5 F 29.7

Min 271.4 F 29.1
Max 273.1 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.4 F 0.0
5:00 231.4 F 0.0
10:00 230.6 F 0.0
15:00 224.8 F 0.0
20:00 222.0 F 0.0
25:00 221.6 F 0.0
30:00 219.6 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 21:23
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:47
CYCLE COMPLETE

03/08/2023 04:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____