

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1352

Name: RITTER2  
Sterilizer ID: V2313228  
07/26/2022 04:21 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 207.3 F 0.0  
2:00 202.1 F 2.4  
4:00 218.0 F 8.4  
6:00 228.9 F 10.0  
8:00 235.5 F 11.1  
10:00 236.4 F 9.5  
12:00 239.0 F 9.3  
14:00 250.4 F 16.0  
16:00 262.8 F 23.1  
18:00 272.3 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.0  
1:00 272.6 F 29.6  
2:00 271.6 F 28.9  
3:00 272.8 F 29.8  
4:00 272.2 F 29.2  
  
Min 271.5 F 28.8  
Max 272.9 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.4 F 0.0  
5:00 227.1 F 0.0  
10:00 237.5 F 0.0  
15:00 225.4 F 0.0  
20:00 228.4 F 0.0  
25:00 225.9 F 0.0  
30:00 225.0 F 0.0  
  
Drying Complete

FILLING: 2:46  
HEATING: 18:11  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:58:09  
CYCLE COMPLETE

07/26/2022 05:19 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_