
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003753
13:39 01/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:07
PRESSURIZING 2:10
STERILIZING 3:09
136.0 C 220kPa 3:09
136.3 C 222kPa 3:39
136.3 C 225kPa 4:09
136.4 C 225kPa 4:39
136.3 C 220kPa 5:09
136.0 C 221kPa 5:39
136.1 C 223kPa 6:09
136.2 C 222kPa 6:39
136.2 C 222kPa 6:39
VENTING 6:39
AIR DRYING 7:25
CYCLE COMPLETE 13:42

Digital Signature #
A28970297A64654D
