
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000659
8:57 20/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:03
PRESSURIZING 3:04
STERILIZING 4:30
136.4 C 223kPa 4:30
136.2 C 221kPa 5:00
136.3 C 224kPa 5:30
136.1 C 221kPa 6:00
136.4 C 223kPa 6:30
136.2 C 220kPa 7:00
136.0 C 216kPa 7:30
136.4 C 222kPa 8:00
136.4 C 222kPa 8:00
VENTING 8:00
AIR DRYING 8:56
CYCLE COMPLETE 8:58

Digital Signature #
27AC0F97BA3B4432
