
Midmark M11 - v1.0.5
Total Cycles: 2186

Name: RITTER2
Sterilizer ID: V2313228
03/28/2023 12:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 135.8 F 0.0
2:00 134.2 F 1.9
4:00 148.8 F 4.8
6:00 175.3 F 9.2
8:00 199.6 F 7.8
10:00 219.5 F 7.5
12:00 232.1 F 12.6
14:00 235.8 F 9.8
16:00 236.2 F 9.0
18:00 248.2 F 14.8
20:00 260.1 F 21.6
22:00 271.2 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.4 F 29.6
2:00 271.3 F 29.1
3:00 272.6 F 29.7
4:00 271.6 F 29.0

Min 271.3 F 28.9
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.8 F 0.0
5:00 214.9 F 0.0
10:00 227.4 F 0.0
15:00 220.9 F 0.0
20:00 220.8 F 0.0
25:00 230.9 F 0.0
30:00 225.4 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 22:40
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:00:59
CYCLE COMPLETE

03/28/2023 01:52 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____