
Midmark M11 - v1.0.0
Total Cycles: 1164

Name: RITTER 1
Sterilizer ID: V2178198
09/13/2021 12:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 122.0 F 0.0
2:00 129.5 F 0.4
4:00 145.0 F 2.7
6:00 170.1 F 6.1
8:00 193.3 F 10.9
10:00 212.7 F 3.9
12:00 226.2 F 8.7
14:00 237.2 F 12.3
16:00 238.6 F 10.6
18:00 235.3 F 8.3
20:00 249.5 F 15.1
22:00 261.2 F 21.7
24:00 272.0 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.6 F 28.9
2:00 271.9 F 28.6
3:00 272.6 F 28.9
4:00 271.7 F 28.4

Min 271.6 F 28.2
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.8 F 0.0
5:00 178.0 F 0.0
10:00 194.7 F 0.0
15:00 193.4 F 0.0
20:00 190.4 F 0.0
25:00 188.4 F 0.0
Dry Cycle Aborted

FILLING: 0:53
HEATING: 24:31
STERILIZING: 4:00
VENTING: 3:07
DRYING: 26:16
TOTAL CYCLE: 00:58:47

09/13/2021 01:39 PM
ACCEPTED: X
REJECTED:
Operator: U01