

-----  
Midmark M11 - v1.0.5  
Total Cycles: 53

Name: RITTER2  
Sterilizer ID: V2313228  
05/27/2021 03:52 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 151.5 F 0.0  
2:00 149.6 F 2.3  
4:00 160.7 F 5.6  
6:00 184.5 F 9.8  
8:00 208.1 F 6.3  
10:00 222.4 F 8.2  
12:00 233.7 F 13.2  
14:00 237.9 F 11.0  
16:00 238.3 F 9.9  
18:00 251.8 F 16.4  
20:00 263.0 F 23.1  
22:00 272.1 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.9  
1:00 271.7 F 28.8  
2:00 272.4 F 29.0  
3:00 271.7 F 29.3  
4:00 271.5 F 29.4

Min 271.1 F 28.8  
Max 272.4 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 176.4 F 0.0  
5:00 232.1 F 0.0  
10:00 226.7 F 0.0  
15:00 209.3 F 0.0  
20:00 202.5 F 0.0  
25:00 202.8 F 0.0  
30:00 198.5 F 0.0  
Drying Complete

FILLING: 1:25  
HEATING: 22:10  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:00:50  
CYCLE COMPLETE

05/27/2021 04:52 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_