
Midmark M11 - v1.0.5
Total Cycles: 3541

Name: RITTER2
Sterilizer ID: V2313228
04/18/2024 10:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.0 F 0.0
2:00 157.3 F 1.9
4:00 171.8 F 5.1
6:00 192.3 F 9.3
8:00 208.2 F 10.0
10:00 220.2 F 8.1
12:00 236.2 F 13.0
14:00 238.3 F 10.1
16:00 236.1 F 8.8
18:00 248.5 F 14.7
20:00 259.8 F 21.2
22:00 270.6 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.4
1:00 272.5 F 29.5
2:00 272.8 F 29.9
3:00 271.7 F 29.6
4:00 272.5 F 29.6

Min 271.4 F 29.3
Max 273.1 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.9 F 0.0
5:00 231.0 F 0.0
10:00 192.9 F 0.0
15:00 187.9 F 0.0
20:00 183.6 F 0.0
25:00 180.7 F 0.0
30:00 179.7 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 23:38
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:02:25
CYCLE COMPLETE

04/18/2024 11:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____