
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001335
10:12 09/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:29
PRESSURIZING 2:33
STERILIZING 3:37
136.0 C 221kPa 3:37
136.4 C 223kPa 4:07
136.4 C 224kPa 4:37
136.4 C 225kPa 5:07
136.4 C 224kPa 5:37
136.4 C 222kPa 6:07
136.3 C 224kPa 6:37
136.0 C 222kPa 7:07
136.0 C 222kPa 7:07
VENTING 7:07
AIR DRYING 8:00
CYCLE COMPLETE 19:31

Digital Signature #
01D11A8965BB849C
