
Midmark M11 - v1.0.0
Total Cycles: 707

Name: MIDMARK
Sterilizer ID: V2178198
02/26/2021 10:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 109.6 F 0.0
2:00 109.7 F 0.0
4:00 127.5 F 0.6
6:00 157.9 F 3.2
8:00 186.2 F 6.9
10:00 215.4 F 11.3
12:00 222.2 F 5.1
14:00 231.3 F 10.2
16:00 231.9 F 7.7
18:00 243.3 F 12.6
20:00 241.2 F 10.7
22:00 254.2 F 17.3
24:00 265.6 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.8 F 29.2
2:00 272.5 F 29.2
3:00 272.1 F 29.0
4:00 271.8 F 28.6

Min 270.8 F 27.9
Max 273.0 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.6 F 0.0
5:00 233.7 F 0.0
10:00 222.7 F 0.0
15:00 222.0 F 0.0
20:00 226.4 F 0.0
25:00 237.8 F 0.0
30:00 228.8 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 25:34
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 01:03:42
CYCLE COMPLETE

02/26/2021 11:22 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____