
Midmark M11 - v1.0.5
Total Cycles: 3431

Name: RITTER2
Sterilizer ID: V2313228
03/04/2024 12:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 94.6 F 0.0
2:00 95.0 F 0.0
4:00 104.4 F 0.2
6:00 131.6 F 2.0
8:00 160.4 F 5.4
10:00 184.9 F 9.3
12:00 203.0 F 13.5
14:00 215.0 F 6.3
16:00 226.3 F 10.6
18:00 230.2 F 7.7
20:00 240.9 F 12.4
22:00 239.7 F 10.7
24:00 252.1 F 16.8
26:00 262.8 F 23.2
28:00 272.8 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.9 F 29.7
2:00 271.7 F 28.9
3:00 272.0 F 29.6
4:00 272.4 F 29.4

Min 271.2 F 28.8
Max 273.1 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.3 F 0.0
5:00 218.2 F 0.0
10:00 204.7 F 0.0
15:00 194.3 F 0.0
20:00 189.3 F 0.0
25:00 186.4 F 0.0
30:00 185.0 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 28:23
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:06:36
CYCLE COMPLETE

03/04/2024 01:12 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____