
Midmark M11 - v1.0.0
Total Cycles: 681

Name: MIDMARK
Sterilizer ID: V2178198
02/19/2021 10:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.3 F 0.0
2:00 140.2 F 1.5
4:00 158.0 F 4.5
6:00 202.0 F 9.6
8:00 220.4 F 4.7
10:00 227.6 F 10.0
12:00 232.3 F 7.5
14:00 243.0 F 12.9
16:00 241.8 F 11.2
18:00 255.4 F 18.1
20:00 267.5 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.6
1:00 271.8 F 28.6
2:00 272.0 F 28.6
3:00 272.5 F 28.8
4:00 272.8 F 29.1

Min 271.7 F 28.5
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.5 F 0.0
5:00 233.1 F 0.0
10:00 225.0 F 0.0
15:00 229.7 F 0.0
20:00 223.0 F 0.0
25:00 212.7 F 0.0
30:00 201.1 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 21:52
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 01:00:01
CYCLE COMPLETE

02/19/2021 11:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____