
Midmark M11 - v1.0.5
Total Cycles: 762

Name: RITTER2
Sterilizer ID: V2313228
02/07/2022 10:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.0 F 0.0
2:00 177.4 F 3.4
4:00 189.5 F 7.8
6:00 209.0 F 13.1
8:00 225.0 F 8.2
10:00 235.1 F 12.1
12:00 240.8 F 11.5
14:00 239.8 F 10.5
16:00 254.1 F 17.7
18:00 266.2 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.8
1:00 272.1 F 29.5
2:00 271.8 F 28.7
3:00 272.3 F 29.5
4:00 272.1 F 28.9

Min 271.2 F 28.6
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.0 F 0.0
5:00 228.2 F 0.0
10:00 218.7 F 0.0
15:00 224.5 F 0.0
20:00 227.4 F 0.0
25:00 227.8 F 0.0
30:00 227.3 F 0.0

Drying Complete

FILLING: 1:30
HEATING: 19:24
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:57:59
CYCLE COMPLETE

02/07/2022 11:55 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____