

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001522  
9:57 23/10/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:32  
PRESSURIZING 2:40  
STERILIZING 3:45  
136.5 C 224kPa 3:45  
136.4 C 226kPa 4:15  
136.3 C 221kPa 4:45  
136.3 C 224kPa 5:15  
136.4 C 222kPa 5:45  
136.4 C 226kPa 6:15  
136.4 C 226kPa 6:45  
136.4 C 223kPa 7:15  
136.4 C 223kPa 7:15  
VENTING 7:16  
AIR DRYING 8:07  
CYCLE COMPLETE 8:11  
-----

Digital Signature #  
DD346C679BB6F3E9  
-----