
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006889
12:06 01/02/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:28
PRESSURIZING 2:31
STERILIZING 3:36
136.4 C 221kPa 3:36
136.3 C 224kPa 4:06
136.2 C 221kPa 4:36
136.4 C 223kPa 5:06
136.3 C 223kPa 5:36
136.1 C 223kPa 6:06
136.3 C 221kPa 6:36
136.4 C 224kPa 7:06
136.4 C 224kPa 7:06
VENTING 7:06
AIR DRYING 7:55
CYCLE COMPLETE 37:55

Digital Signature #
0D23B3A8659CCD1D
