
Midmark M11 - v1.0.0
Total Cycles: 2897

Name: RITTER 1
Sterilizer ID: V2178198
02/14/2023 11:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 211.9 F 0.0
2:00 198.5 F 4.4
4:00 203.7 F 9.3
6:00 220.8 F 7.3
8:00 229.5 F 9.4
10:00 236.9 F 9.1
12:00 242.8 F 12.4
14:00 241.0 F 10.8
16:00 254.3 F 17.5
18:00 265.6 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.3 F 29.0
2:00 271.5 F 28.3
3:00 272.0 F 29.1
4:00 272.0 F 28.8

Min 271.2 F 28.2
Max 272.4 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.5 F 0.0
5:00 234.1 F 0.0
10:00 226.5 F 0.0
15:00 229.0 F 0.0
20:00 229.2 F 0.0
25:00 229.0 F 0.0
30:00 230.3 F 0.0

Drying Complete

FILLING: 1:49
HEATING: 19:45
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:48
CYCLE COMPLETE

02/14/2023 12:07 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____