
Midmark M11 - v1.0.5
Total Cycles: 965

Name: RITTER2
Sterilizer ID: V2313228
04/07/2022 06:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.1 F 0.0
2:00 81.9 F 0.0
4:00 84.8 F 0.0
6:00 109.6 F 0.0
8:00 155.8 F 1.5
10:00 163.4 F 5.3
12:00 188.8 F 9.0
14:00 209.2 F 12.9
16:00 223.8 F 6.5
18:00 235.8 F 10.6
20:00 233.8 F 8.1
22:00 242.5 F 12.5
24:00 240.5 F 11.0
26:00 253.0 F 17.2
28:00 263.8 F 23.7
30:00 271.8 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.4
1:00 272.0 F 29.3
2:00 272.4 F 29.3
3:00 272.8 F 29.5
4:00 272.8 F 29.7

Min 271.5 F 29.2
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.7 F 0.0
5:00 219.2 F 0.0
10:00 190.5 F 0.0
15:00 189.8 F 0.0
20:00 215.8 F 0.0
25:00 221.3 F 0.0
30:00 230.8 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 30:48
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:08:53
CYCLE COMPLETE

04/07/2022 08:07 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____