
Midmark M11 - v1.0.0
Total Cycles: 2613

Name: RITTER 1
Sterilizer ID: V2178198
11/28/2022 01:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.4 F 0.0
2:00 201.3 F 4.7
4:00 210.5 F 10.1
6:00 223.9 F 6.2
8:00 236.0 F 11.7
10:00 236.4 F 9.4
12:00 237.8 F 9.0
14:00 249.2 F 15.1
16:00 261.4 F 22.1
18:00 272.6 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.6 F 29.2
2:00 272.0 F 28.9
3:00 272.5 F 29.1
4:00 271.8 F 28.7

Min 271.7 F 28.5
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.7 F 0.0
5:00 230.9 F 0.0
10:00 231.1 F 0.0
15:00 229.6 F 0.0
20:00 226.0 F 0.0
25:00 224.3 F 0.0
30:00 224.3 F 0.0
Drying Complete

FILLING: 3:03
HEATING: 18:23
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:39
CYCLE COMPLETE

11/28/2022 02:21 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____