

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2288

Name: RITTER 1  
Sterilizer ID: V2178198  
09/06/2022 07:14 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.5 F 0.0  
2:00 73.8 F 0.0  
4:00 80.6 F 0.2  
6:00 108.8 F 0.7  
8:00 140.7 F 1.5  
10:00 173.9 F 2.8  
12:00 193.5 F 6.1  
14:00 215.2 F 9.9  
16:00 225.2 F 10.6  
18:00 228.1 F 7.3  
20:00 238.7 F 11.6  
22:00 233.9 F 8.1  
24:00 245.7 F 13.0  
26:00 243.6 F 12.0  
28:00 256.0 F 18.4  
30:00 266.8 F 25.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.2  
1:00 272.6 F 28.9  
2:00 271.5 F 28.1  
3:00 272.6 F 29.0  
4:00 272.0 F 28.3

Min 271.5 F 28.0  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.5 F 0.0  
5:00 231.2 F 0.0  
10:00 222.1 F 0.0  
15:00 178.6 F 0.0  
20:00 177.1 F 0.0  
25:00 169.7 F 0.0  
30:00 166.0 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 31:34  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:09:41  
CYCLE COMPLETE

09/06/2022 08:23 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_