

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000888  
13:08 27/01/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:20  
PRESSURIZING 4:21  
STERILIZING 5:36  
136.0 C 220kPa 5:36  
136.0 C 230kPa 6:06  
136.2 C 230kPa 6:36  
136.4 C 233kPa 7:06  
136.4 C 233kPa 7:36  
136.4 C 232kPa 8:06  
136.3 C 232kPa 8:36  
136.4 C 232kPa 9:06  
136.4 C 232kPa 9:06  
VENTING 9:06  
AIR DRYING 10:04  
CYCLE COMPLETE 25:04  
-----

Digital Signature #  
CCDB116E7ADC3FA0  
-----