

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3673

Name: RITTER2  
Sterilizer ID: V2313228  
06/06/2024 11:14 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 89.9 F 0.0  
2:00 89.9 F 0.2  
4:00 96.6 F 1.5  
6:00 122.8 F 3.8  
8:00 152.6 F 6.9  
10:00 172.9 F 10.5  
12:00 204.9 F 6.7  
14:00 213.5 F 7.0  
16:00 226.5 F 10.9  
18:00 228.9 F 7.6  
20:00 237.2 F 12.1  
22:00 237.1 F 10.3  
24:00 249.2 F 16.0  
26:00 259.7 F 22.0  
28:00 269.7 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 30.4  
1:00 272.1 F 30.7  
2:00 272.2 F 30.7  
3:00 272.8 F 30.6  
4:00 272.7 F 30.5

Min 271.9 F 30.3  
Max 272.9 F 30.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.0 F 0.0  
5:00 222.5 F 0.0  
10:00 213.8 F 0.0  
15:00 204.4 F 0.0  
20:00 205.2 F 0.0  
25:00 218.3 F 0.0  
30:00 220.9 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 29:44  
STERILIZING: 4:00  
VENTING: 3:39  
DRYING: 30:00  
TOTAL CYCLE: 01:08:14  
CYCLE COMPLETE

06/06/2024 12:22 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_