
Midmark M11 - v1.0.0
Total Cycles: 892

Name: RITTER 1
Sterilizer ID: V2178198
06/14/2021 10:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.9 F 0.0
2:00 149.4 F 1.4
4:00 154.5 F 4.0
6:00 181.7 F 7.9
8:00 211.7 F 12.5
10:00 219.2 F 6.2
12:00 230.7 F 11.0
14:00 233.8 F 7.3
16:00 243.9 F 13.3
18:00 243.3 F 11.9
20:00 255.3 F 18.3
22:00 266.5 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.2 F 28.7
2:00 273.0 F 29.2
3:00 271.9 F 28.8
4:00 272.4 F 28.8

Min 271.6 F 28.4
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.3 F 0.0
5:00 200.7 F 0.0
10:00 199.7 F 0.0
15:00 202.1 F 0.0
20:00 205.0 F 0.0
25:00 207.5 F 0.0
Dry Cycle Aborted

FILLING: 1:05
HEATING: 23:31
STERILIZING: 4:00
VENTING: 3:14
DRYING: 29:03
TOTAL CYCLE: 01:00:54

06/14/2021 11:53 AM
ACCEPTED: X
REJECTED:
Operator: U01