
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000746
12:45 23/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:29
PRESSURIZING 2:38
STERILIZING 3:49
136.5 C 233kPa 3:49
136.1 C 231kPa 4:19
136.4 C 233kPa 4:49
136.2 C 231kPa 5:19
136.4 C 231kPa 5:49
136.3 C 233kPa 6:19
136.3 C 233kPa 6:49
136.0 C 231kPa 7:19
136.0 C 231kPa 7:19
VENTING 7:19
AIR DRYING 8:10
CYCLE COMPLETE 9:45

Digital Signature #
2700C70A727A67CD
