

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 005484  
14:57 03/05/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 0:50  
PRESSURIZING 0:55  
STERILIZING 2:03  
136.4 C 223kPa 2:03  
136.4 C 225kPa 2:33  
136.1 C 222kPa 3:03  
136.4 C 224kPa 3:33  
136.4 C 224kPa 4:03  
136.3 C 224kPa 4:33  
136.4 C 225kPa 5:03  
136.4 C 225kPa 5:33  
136.4 C 225kPa 5:33  
VENTING 5:33  
AIR DRYING 6:21  
CYCLE COMPLETE 6:25

-----  
Digital Signature #  
F61C3D03C5FBDC60  
-----