
Midmark M11 - v1.0.0
Total Cycles: 842

Name: RITTER 1
Sterilizer ID: V2178198
05/27/2021 09:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.4 F 0.0
2:00 179.5 F 2.5
4:00 181.2 F 6.2
6:00 203.0 F 10.9
8:00 220.1 F 5.4
10:00 225.6 F 10.8
12:00 230.0 F 7.1
14:00 243.0 F 13.5
16:00 242.9 F 11.9
18:00 255.5 F 18.4
20:00 266.5 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.5 F 29.4
2:00 271.9 F 28.8
3:00 271.4 F 28.8
4:00 272.5 F 29.3

Min 271.2 F 28.5
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.1 F 0.0
5:00 205.0 F 0.0
10:00 201.9 F 0.0
15:00 204.9 F 0.0
20:00 233.4 F 0.0
25:00 225.7 F 0.0
30:00 218.4 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 21:37
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:00:17
CYCLE COMPLETE

05/27/2021 10:38 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____